

University of Groningen

Magnetic resonance imaging in patients with cardiac implantable electronic devices

Maass, A H; Hemels, M E W; Allaart, C P

Published in:
Netherlands Heart Journal

DOI:
[10.1007/s12471-018-1192-3](https://doi.org/10.1007/s12471-018-1192-3)

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version
Publisher's PDF, also known as Version of record

Publication date:
2018

[Link to publication in University of Groningen/UMCG research database](#)

Citation for published version (APA):

Maass, A. H., Hemels, M. E. W., & Allaart, C. P. (2018). Magnetic resonance imaging in patients with cardiac implantable electronic devices. *Netherlands Heart Journal*, 26(12), 584-590.
<https://doi.org/10.1007/s12471-018-1192-3>

Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.



Magnetic resonance imaging in patients with cardiac implantable electronic devices

A. H. Maass¹ · M. E. W. Hemels^{2,3} · C. P. Allaart⁴

© The Author(s) 2018

Abstract

In recent years the prevalence of implantation of a cardiac implantable electronic device (CIED) has increased due to expanding implantation indications and prolonged life expectancy. Diagnostic strategies increasingly employ magnetic resonance imaging (MRI) to aid therapeutic strategies. In earlier guidelines, MRI was contra-indicated in patients with CIEDs, mainly due to previous reports of severe complications. With the development of MRI-conditional CIEDs and recent evidence concerning non-MRI-conditional CIEDs, MRIs in CIED patients can be safely performed in many hospitals. However, there are several questions that need to be addressed. Which patients can we scan? How can the scans be performed safely? And last but not least, can cardiac MRI provide diagnostic yield in patients with CIEDs? Current European guidelines are rather outdated and vague about patient selection and practical issues. There are national guidelines on this topic but several issues need extra attention and those are addressed in this point of view. It is important to create an environment with proper patient selection without unnecessary MRI scans in CIED patients, but also without unnecessary fear of complications, preventing access to MRI in patients who can benefit from this powerful diagnostic tool.

Keywords MRI · Pacemaker · ICD · Safety

Introduction

In recent years the prevalence of implantation of a cardiac implantable electronic device (CIED) has increased [1] due to expanding implantation indications and prolonged life expectancy. This increase has been observed with both pacemakers and implantable cardioverter defibrillators (ICDs), as well as with implantable loop recorders. As pacemaker and ICD patients are prone to develop significant comorbidities, it is crucial that protocols for medical

diagnostic and therapeutic procedures are established and followed to avoid temporary or permanent device malfunctioning (reviewed in [2]).

Diagnostic strategies increasingly employ magnetic resonance imaging (MRI). Over the last decades the number of MRI scans has doubled every 5 years [3]. It is estimated that 28% of CIED patients will have an indication for MRI scanning in a period of 4 years with one third of patients needing more than one scan [3]. In earlier guidelines, MRI was contra-indicated in patients with a CIED, mainly due to previous reports of severe complications. Major safety concerns were induction of arrhythmias, tip heating, mechanical concerns and electromagnetic effects [2].

With the development of MRI-conditional pacemakers in 2008, MRI-conditional ICDs in 2011, and recent evidence concerning non-MRI-conditional CIEDs, performing MRI scans in CIED patients has become part of daily routine in many hospitals. Whereas first-generation MRI-conditional devices had specially developed leads with a higher chance of peri-operative complications [4], presently most current leads and devices have an MRI-conditional label. In the latest European pacing guidelines, performing MRI in patients with MRI-conditional devices has received a Class IIa indication whereas non-MRI-conditional devices have

✉ A. H. Maass
a.h.maass@umcg.nl

¹ Department of Cardiology, Thorax center, University of Groningen, University Medical Center Groningen, Groningen, The Netherlands

² Department of Cardiology, Rijnstate Hospital, Arnhem, The Netherlands

³ Radboud University Hospital, University of Nijmegen, Nijmegen, The Netherlands

⁴ Cardiology, Amsterdam Cardiovascular Sciences, Amsterdam UMC, Vrije Universiteit Amsterdam, Amsterdam, The Netherlands



received a Class IIb indication for MRI scanning [5]. Since the publication of these, rather outdated, guidelines several new publications increased the body of evidence concerning safety of MRI in patients with non-MRI-conditional devices. A fairly recent expert consensus statement of the Heart Rhythm Society provides extensive recommendations on several issues [6]. The Netherlands Society of Cardiology has also published practice guidelines surrounding MRI and CIEDs (<https://www.nvvc.nl/media/richtlijn/237/Leidraad%20MRI.pdf>).

A recent study confirms that the accumulating evidence of safe MRI in CIED patients has not yet translated into wide-spread access to MRI. Still, in patients with pacemakers, and even more so in patients with ICDs, MRI is denied by cardiologists or radiologists [7].

This point of view focuses on several practical issues: patient selection, considerations for programming and patient safety. Patients who might benefit from MRI should preferably not be denied access to this powerful diagnostic tool. We also discuss the possibility of performing cardiac MRI in patients with CIEDs as this is thought to be difficult, if not impossible, due to artefacts created by the metal-containing parts of the CIED system.

Patient selection

Modern CIEDs from most manufacturers are labelled MRI-conditional. This concerns both pacemakers and ICDs, and includes modern techniques such as leadless pacemakers [8] and the subcutaneous ICD [9]. Several electrodes that have been on the market for quite some time also received the MRI-conditional label and when they are connected to a modern pacemaker or ICD from the same manufacturer, the whole system is MRI-conditional. Frequently, during device replacement or *de novo* implantations MRI-conditional components of several manufacturers are mixed and in that case the whole system cannot be labelled MRI-conditional.

MRI-conditional labelling is usually acquired by technical construction calculations and *in vitro* testing. Clinical patient data before achievement of the label were scarce but are accumulating after introduction into the market. For FDA licensing purposes, 2,629 CIED patients were prospectively followed and underwent a total of 872 MRI scans without any complications [3]. Clinical data are available in non-MRI-conditional systems as well. Currently, these data are robust with a very large prospective registry including 1,500 patients who underwent non-thoracic MRI in the MagnaSafe registry [10] and a recent meta-analysis which included 5,099 patients [11]. As yet, less data are available on the safety of MRI in patients with abandoned leads or epicardial leads. Even though several

studies showed no serious adverse events in patients with abandoned leads [12, 13], sensitive markers of lead or tissue damage, such as changes in impedance or threshold, cannot be assessed. Other investigators found no problems with several issues posing a contraindication for MRI, such as batteries close to depletion and recall components [14]. Epicardial electrodes might pose a greater risk for use in the MRI environment since they have the disadvantage of not being cooled by blood flow. They were typically used in children and patients with congenital heart disease, but are nowadays increasingly used in patients who need cardiac resynchronisation therapy as well [15]. The lifetime chance of the need for MRI in young patients especially is very high and we urgently need data demonstrating the safety of MRI in patients with epicardial electrodes. The same holds true for patients with older, non-MRI-conditional coronary sinus electrodes.

Practical issues around the MRI scan

To assure that a patient can be scanned safely, several steps have to be taken. These practical considerations evolve around the CIED system, but also involve MRI settings and patient monitoring during the scan (summarized in Fig. 1).

MRI scans in CIED patients require specific attention from the referring clinical department as well as the radiology and the cardiology departments. Therefore, as a first step, these labour-intensive scans should be reserved for patients that actually derive benefit, i.e. the scan should have clear therapeutic or prognostic implications that cannot be obtained by other diagnostic means. Before scheduling the patient this needs to be taken into account. Secondly, it needs to be determined whether the patient has an MRI-conditional or non-MRI-conditional system. According to manufacturer specifications of the MRI-conditional system, an MRI technical specialist should document the specific MRI conditions (e.g. Tesla, specific absorption rates, etcetera) (see Table 1). Thirdly, information has to be available about pacemaker dependency and recent technical measurements. Measurements should be within normal range and devices should not be near the end of their battery life given the chance of switching to a safe replacement mode during the scan.

At the day of the scan, actual technical measurements should be performed and the device has to be reprogrammed according to manufacturer's specifications in MRI-conditional devices. This specific programming is part of the conditions under which MRI is possible. In pacemaker-dependent patients an asynchronous mode should be programmed (VOO or DOO). In non-pacemaker-dependent patients a VVI or DDI mode should be programmed, however, in patients with high pacing percentages asynchronous pac-

Fig. 1 Flow diagram of patient selection and device programming for magnetic resonance imaging. *MRI* magnetic resonance imaging, *HR* heart rate, *PM* pacemaker, *ICD* implantable cardioverter defibrillator, *CS* coronary sinus, *ECG* electrocardiogram




Table 1 Technical considerations for MRI in patients with cardiac implantable electronic devices

Technical specification of magnetic resonance imaging

For MRI-conditional devices follow manufacturer's instructions for scan

For non-MRI-conditional devices:

- 1.5 T MRI should be preferred
- Gradient slew rate should not exceed 200 T/m/s
- Dorsal patient position
- Imaging landmark near the device (thorax) should be avoided
- Local transmit coils should be avoided
- SAR and scan time should be limited

Emergency equipment/external defibrillator as well as a device programmer should be present during the MRI

Continuous patient monitoring (electrocardiography/pulse oximetry) during the MRI should be performed

MRI magnetic resonance imaging, *SAR* specific absorption rate

ing modes might be preferable. Additional advanced pacing features should be deactivated. In ICD patients all ICD therapies should be deactivated (see Table 2).

During the scan the patient should be monitored using electrocardiography (ECG) and/or plethysmography. The latter may be the preferable technique because ECG monitoring might be disturbed during the scan, whereas plethysmographic signals usually are not. However, plethysmography obviously remains sensitive to loss of contact if the patient is moving his fingers. Therefore, monitoring should be performed by a professional trained in haemodynamic mon-

itoring who is also aware of the possible scenarios needing intervention (Table 3). 'Power-on resets' occur relatively frequently, especially in non-MRI-conditional devices leading to a change of pacing mode that is susceptible to oversensing [12]. An external defibrillator with external pacing capability should always be present in close proximity of the MRI scanner and the available personnel needs to be trained in its use. Currently, there is increasing evidence that magnet resonance scanning can also be performed without monitoring, especially in acute settings when monitoring is not readily available, due to the low rate of problems during the scan [16].

After the scan, the programme settings of the device should be restored to the pre-MRI settings and technical measurements should be performed. Special attention should be paid to changes in electrode impedances, sensing, pacing thresholds and battery measurements. In case of significant changes a prolonged period of observation should be considered.

Cardiac MRI in patients with CIEDs

Non-invasive cardiac imaging is important for diagnostics [17] and planning of therapeutic interventions [18].

Previously, cardiac MRI was thought to be difficult, if not impossible, in patients with CIEDs due to artefacts created by the metal-containing parts of the CIED system.

Table 2 Advice for programming devices in non-MRI-conditional patients undergoing magnetic resonance scanning

In pacemaker-dependent patients: asynchronous mode (VOO/DOO) with HF of 80–90/min with high output, RV-only pacing in biventricular devices

In non-pacemaker-dependent patients:

- Resting heart rate >50/min: VVI or DDI 30–40/min
- Resting heart rate <50/min: consider VOO/DOO 80–90/min with high output
- RV-only pacing in CRT patients

Deactivate additional features:

- Rate response mode
- Anti-tachycardia therapies (including anti-tachycardia pacing and shocks)—ICD only
- LV-triggered pacing (ventricular sense response)—biventricular devices only
- Anti-pacemaker-mediated tachycardia pacing (PMT algorithms)
- PVC-triggered pacing response
- PAC-triggered pacing response
- Atrial fibrillation therapies (rate smoothing, overdrive pacing, conducted atrial fibrillation response)
- Hysteresis pacing
- Magnet response (if the option exists)
- Noise response

MRI magnetic resonance imaging, *HF* heart frequency, *RV* right ventricular, *CRT* cardiac resynchronisation therapy, *ICD* implantable cardioverter defibrillator, *LV* left ventricular, *PMT* pacemaker-mediated tachycardia, *PVC* premature ventricular contraction, *PAC* premature atrial contraction

Table 3 Scenarios during magnetic resonance imaging requiring immediate action of the observing professional

Scenario	Possible reasons
Asystole	Device malfunction in asynchronous mode 'Power-on reset' with pacing inhibition by noise True bradycardia in non-pacemaker dependent patients in inhibitory mode
Ventricular fibrillation/tachycardia	R-on-T phenomenon in asynchronous pacing mode True arrhythmia
Sudden rise in heart rate	'Power-on reset' in inhibitory mode leading to pacing True endogenous tachycardia
Sudden drop in heart rate	'Power-on reset' in asynchronous mode True endogenous bradycardia in inhibitory mode

Recent developments have shown that with specific MRI settings cardiac MRIs can be performed with reasonable quality. This is especially true in pacemaker implants in the right pectoral region and to a lesser degree in left-sided pacemakers and even less in ICDs due to the larger metal-containing battery. In non-MRI-conditional systems, thoracic MRIs have only partly been included in the large prospective registries but seem to be safe [19]. In this study, wideband pulse sequence for late gadolinium enhancement yields a high rate of studies unaffected by artefacts.

Conclusions

MRI can be safely performed in both patients with MRI-conditional and non-MRI-conditional systems if specific programming is performed, technical limitations of magnetic resonance scanning are taken into consideration and patients are adequately monitored. Even cardiac MRI can be performed if the correct protocols are followed. Cardiologists and radiologists should be familiar with the protocols and not deny patients access to this powerful diagnostic tool.

Conflict of interest A.H. Maass received lecture fees from Medtronic and LivaNova. M.E.W. Hemels received lecture fees from Biotronik and LivaNova. C.P. Allaart received research grants from Boston Scientific and Abbott and lecture fees from Biotronik.

Open Access This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made.

References

- van Veldhuisen DJ, Maass AH, Priori SG, et al. Implementation of device therapy (cardiac resynchronization therapy and implantable cardioverter defibrillator) for patients with heart failure in Europe: changes from 2004 to 2008. *Eur J Heart Fail*. 2009;11:1143–51.
- Nordbeck P, Ertl G, Ritter O. Magnetic resonance imaging safety in pacemaker and implantable cardioverter defibrillator patients: how far have we come? *Eur Heart J*. 2015;36:1505–11.
- Williamson BD, Gohn DC, Ramza BM, et al. Real-World Evaluation of Magnetic Resonance Imaging in Patients With a Magnetic Resonance Imaging Conditional Pacemaker System: Results of 4-Year Prospective Follow-Up in 2,629 Patients. *Jacc Clin Electrophysiol*. 2017;3:1231–9.
- Shurrah M, Kaoutskaia A, Baranchuk A, et al. Are there increased periprocedural complications with the MRI-conditional Medtronic Revo SureScan Pacing System? : A meta-analysis. *Neth Heart J*. 2018;26:233–9.
- European Society of C, European Heart Rhythm A, Brignole M, Auricchio A, Baron-Esquivias G, et al. 2013 ESC guidelines on cardiac pacing and cardiac resynchronization therapy: the task force on cardiac pacing and resynchronization therapy of the European Society of Cardiology (ESC). Developed in collaboration with the European Heart Rhythm Association (EHRA). *Europace*. 2013;15:1070–118.
- Indik JH, Gimbel JR, Abe H, et al. 2017 HRS expert consensus statement on magnetic resonance imaging and radiation exposure in patients with cardiovascular implantable electronic devices. *Heart Rhythm*. 2017;14:e97–e153.
- Celentano E, Caccavo V, Santamaria M, et al. Access to magnetic resonance imaging of patients with magnetic resonance-conditional pacemaker and implantable cardioverter-defibrillator systems: results from the Really ProMRI study. *Europace*. 2018;20:1001–9.
- Soejima K, Edmonson J, Ellingson ML, et al. Safety evaluation of a leadless transcatheter pacemaker for magnetic resonance imaging use. *Heart Rhythm*. 2016;13:2056–63.



9. Keller J, Neuzil P, Vymazal J, et al. Magnetic resonance imaging in patients with a subcutaneous implantable cardioverter-defibrillator. *Europace*. 2015;17:761–6.
10. Russo RJ, Costa HS, Silva PD, et al. Assessing the Risks Associated with MRI in Patients with a Pacemaker or Defibrillator. *N Engl J Med*. 2017;376:755–64.
11. Shah AD, Morris MA, Hirsh DS, et al. Magnetic resonance imaging safety in nonconditional pacemaker and defibrillator recipients: A meta-analysis and systematic review. *Heart Rhythm*. 2018;15:1001–8.
12. Higgins JV, Sheldon SH, Watson RE Jr., et al. “Power-on resets” in cardiac implantable electronic devices during magnetic resonance imaging. *Heart Rhythm*. 2015;12:540–4.
13. Padmanabhan D, Kella DK, Mehta R, et al. Safety of magnetic resonance imaging in patients with legacy pacemakers and defibrillators and abandoned leads. *Heart Rhythm*. 2018;15:228–33.
14. Horwood L, Attili A, Luba F, et al. Magnetic resonance imaging in patients with cardiac implanted electronic devices: focus on contraindications to magnetic resonance imaging protocols. *Europace*. 2017;19:812–7.
15. Schouwenburg JJ, Klinkenberg TJ, Maass AH, Mariani MA. Video-assisted thoracic placement of epicardial leads. *J Card Surg*. 2014;29:286–9.
16. Bertelsen L, Petersen HH, Philbert BT, et al. Safety of magnetic resonance scanning without monitoring of patients with pacemakers. *Europace*. 2017;19:818–23.
17. Baroni M, Nava S, Quattrocchi G, et al. Role of cardiovascular magnetic resonance in suspected cardiac amyloidosis: late gadolinium enhancement pattern as mortality predictor. *Neth Heart J*. 2018;26:34–40.
18. Maass AH, Yap SC. Imaging before cardiac resynchronisation therapy implantation—luxury or necessity? *Neth Heart J*. 2018;26:422–4.
19. Do DH, Eyvazian V, Bayoneta AJ, et al. Cardiac magnetic resonance imaging using wideband sequences in patients with nonconditional cardiac implanted electronic devices. *Heart Rhythm*. 2018;15:218–25.